

Calcium Chloride, Anhydrous
BAKER ANALYZED® A.C.S. Reagent
20 mesh and finer, for desiccant use



Material No.: 1311-05
Batch No.: 0000000036
Manufactured Date: 2012/03/13

Certificate of Analysis

Meets ACS Reagent Chemical Requirements,

Test	Specification	Result
Assay (CaCl ₂) (by Ag titrn)	≥ 96.0 %	99.3
Titration Base (meq/g)	≤ 0.006	< 0.001
Mesh - On U.S. No. 20 Sieve	≤ 10 %	3
Mesh - Thru U.S. No. 200 Sieve	≤ 30 %	5

For Laboratory, Research, or Manufacturing Use

Storage Conditions:

Country of Origin: J P

Packaging Site: Paris Mfg Ctr & DC

ISO

Phillipsburg, NJ 9001.2008, 14001.2004
Paris, KY 9001.2008
Mexico city, Mexico 9001.2008
Deventer, The Netherlands 9001.2008, 14001.2004
Selangor, Malaysia 9001.2008
Panoli, India 9001.2008
Gliwice, Poland 9001.2008, 17025.2005

Richard M Siberski
Global Director of Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600
Avantor™ Performance Materials Inc.
3477 Corporate Parkway, Suite #200, Center Valley, PA 18034, U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610