

Methanol
BAKER ANALYZED® LC-MS Reagent
For Use in Liquid Chromatography and Mass Spectrometry



Material No.: 9830-02
Batch No.: 0000027434
Manufactured Date: 2012/10/17
Expiration Date: 2014/10/17

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	>= 99.9 %	100.0
Appearance	Passes Test	PT
Gradient Elution Test (a.u.) – 254 nm	<= 0.01	< 0.01
Fluorescence Trace Impurities, measured as Quinine Base – at 450 nm Emission	<= 0.30 ppb	< 0.1
Fluorescence Trace Impurities, measured as Quinine Base – at Emission Maximum for Impurities	<= 1.0 ppb	0.1
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.1
Water (H ₂ O)	<= 500 ppm	< 50
LC/MS Suitability – Largest Response on ESI- Positive Mode (as Reserpine)	<= 50 ppb	< 50
Trace Impurities – Aluminum (Al)	<= 50.0 ppb	< 5.0
Trace Impurities – * Calcium (Ca)	<= 50.0 ppb	2.0
Trace Impurities – Iron (Fe)	<= 50.0 ppb	< 1.0
Trace Impurities – Lithium (Li)	<= 30.0 ppb	< 5.0
Trace Impurities – Magnesium (Mg)	<= 50.0 ppb	< 1.0
Trace Impurities – Nickel (Ni)	<= 30.0 ppb	< 5.0
Trace Impurities – Potassium (K)	<= 50.0 ppb	< 10.0
Trace Impurities – * Sodium (Na)	<= 50.0 ppb	< 5.0

For Laboratory, Research or Manufacturing Use

* May change over time due to extraction from glass container.

Filtered through a 0.2 micron filter.

Material No.: 9830-02
Batch No.: 0000027434

Country of Origin: US
Packaging Site: Phillipsburg Mfg Ctr & DC



Phillipsburg, NJ 9001:2008, 14001:2004
Paris, KY 9001:2008
Mexico City, Mexico 9001:2008
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003
Gliwice, Poland 9001:2008, 17025:2005
Selangor, Malaysia 9001:2008
Dehradun, India, 9001:2008, 14001:2004, 13485:2003
Mumbai, India, 9001:2008, 17025:2005
Panoli, India 9001:2008

A handwritten signature in white ink, appearing to read 'Richard M Siberski', is written over the blue background.

Richard M Siberski
Global Director of Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600
Avantor™ Performance Materials Inc.
3477 Corporate Parkway, Suite #200, Center Valley, PA 18034, U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610