

Methanol

HPLC

For use in Liquid Chromatography (HPLC & UHPLC) &

Spectrophotometry

(methyl alcohol)



Material No.: 9093-03

Batch No.: 0000025528

Manufactured Date: 2012/10/27

Retest Date: 2017/10/26

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	>= 99.8 %	100.0
Ultraviolet Absorbance (1.00-cm cell vs. water) - 400-254 nm	<= 0.01	0.01
Ultraviolet Absorbance (1.00-cm cell vs. water) - 225 nm	<= 0.15	0.09
Ultraviolet Absorbance (1.00-cm cell vs. water) - UV Cut-off, nm	<= 205	< 203
Gradient Elution Test (a.u.) - 254 nm	<= 0.002	< 0.001
Fluorescence Trace Impurities, measured as Quinine Base - at 450 nm Emission	<= 0.3 ppb	0.1
Fluorescence Trace Impurities, measured as Quinine Base - at Emission Maximum for Impurities	<= 1.0 ppb	0.2
Acetone	<= 0.001 %	< 0.001
Residue after Evaporation	<= 1.0000 ppm	< 0.1000
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.1	< 0.00
Water (by KF, coulometric)	<= 0.05 %	< 0.01

For Laboratory, Research or Manufacturing Use

Country of Origin: US

Packaging Site: Phillipsburg Mfg Ctr & DC



Phillipsburg, NJ 9001:2008, 14001:2004
Paris, KY 9001:2008
Mexico City, Mexico 9001:2008
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003
Gliwice, Poland 9001:2008, 17025:2005
Selangor, Malaysia 9001:2008
Dehradun, India, 9001:2008, 14001:2004, 13485:2003
Mumbai, India, 9001:2008, 17025:2005
Panoli, India 9001:2008

Richard M Siberski
Global Director of Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600
Avantor™ Performance Materials Inc.

3477 Corporate Parkway, Suite #200, Center Valley, PA 18034, U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610