

Acetone

HPLC

For use in Liquid Chromatography (HPLC & UHPLC) &
Spectrophotometry



Material No.: 9002-33
Batch No.: 0000028880
Manufactured Date: 2012/10/18
Retest Date: 2017/10/17

Certificate of Analysis

Test	Specification	Result
Ultraviolet Absorbance (1.00-cm cell vs. water) - 400-350 nm	<= 0.01	<0.01
Ultraviolet Absorbance (1.00-cm cell vs. water) - UV Cut-off, nm	<= 330	< 329
Fluorescence Trace Impurities, measured as Quinine Base - at 450 nm Emission	<= 0.3 ppb	<0.1
Fluorescence Trace Impurities, measured as Quinine Base - at Emission Maximum for Impurities	<= 1.0 ppb	<0.1
Assay ((CH ₃) ₂ CO)(by GC)	>= 99.3 %	100.0
Residue after Evaporation	<= 2.0 ppm	< 1.0
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.4 %	0.2

For Laboratory, Research or Manufacturing Use
Filtered through a 0.2 micron filter.
Packaged under Nitrogen.

Country of Origin: US
Packaging Site: Phillipsburg Mfg Ctr & DC



Phillipsburg, NJ 9001:2008, 14001:2004
Paris, KY 9001:2008
Mexico City, Mexico 9001:2008
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003
Gliwice, Poland 9001:2008, 17025:2005
Selangor, Malaysia 9001:2008
Dehradun, India, 9001:2008, 14001:2004, 13485:2003
Mumbai, India, 9001:2008, 17025:2005
Panoli, India 9001:2008

Richard M Siberski
Global Director of Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600
Avantor™ Performance Materials Inc.

3477 Corporate Parkway, Suite #200, Center Valley, PA 18034. U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610