

Ammonium Acetate
HPLC
For Use in High Performance Liquid Chromatography



Material No.: 0599-08
Batch No.: 0000070926
Manufactured Date: 2012/01/25
Retest Date: 2019/01/23

Certificate of Analysis

| Test | Specification | Result |
|--|----------------------|-----------|
| Assay ($\text{CH}_3\text{COONH}_4$) (by acid-base titration) | $\geq 97 \%$ | 99 |
| ACS – Insoluble Matter | $\leq 0.005 \%$ | 0.001 |
| Residue after Ignition | $\leq 0.010 \%$ | 0.002 |
| pH of 5% Solution at 25°C | 6.7 – 7.3 | 6.9 |
| Nitrate (NO_3) | $\leq 0.001 \%$ | < 0.001 |
| Sulfate (SO_4) | $\leq 0.001 \%$ | < 0.001 |
| UV Absorbance (1 M Aqueous Solution) (1.00-cm cell vs. water) – 350 nm | ≤ 0.010 | < 0.002 |
| UV Absorbance (1 M Aqueous Solution) (1.00-cm cell vs. water) – 280 nm | ≤ 0.010 | 0.002 |
| UV Absorbance (1 M Aqueous Solution) (1.00-cm cell vs. water) – 254 nm | ≤ 0.020 | 0.004 |
| Chloride (Cl) | $\leq 5 \text{ ppm}$ | 5 |
| Trace Impurities – ACS – Heavy Metals (as Pb) | $\leq 5 \text{ ppm}$ | < 3 |
| Trace Impurities – Iron (Fe) | $\leq 5 \text{ ppm}$ | < 3 |

For Laboratory, Research or Manufacturing Use

Country of Origin: US
Packaging Site: Paris Mfg Ctr & DC



Phillipsburg, NJ 9001:2008, 14001:2004
Paris, KY 9001:2008
Mexico City, Mexico 9001:2008
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003
Gliwice, Poland 9001:2008, 17025:2005
Selangor, Malaysia 9001:2008
Dehradun, India, 9001:2008, 14001:2004, 13485:2003
Mumbai, India, 9001:2008, 17025:2005
Panoli, India 9001:2008

Richard M Siberski
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600
Avantor™ Performance Materials Inc.

3477 Corporate Parkway, Suite #200, Center Valley, PA 18034, U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610