

Acetone
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9254-03
Batch No.: 0000083802
Manufactured Date: 2014/05/30
Expiration Date: 2017/05/29

Certificate of Analysis

| Test | Specification | Result |
|---|---------------|--------|
| Assay ((CH ₃) ₂ CO) (by GC, corrected for water) | >= 99.4 % | 99.6 |
| Color (APHA) | <= 10 | 10 |
| Residue after Evaporation | <= 1.0000 ppm | 0.9000 |
| Substances Reducing Permanganate | Passes Test | PT |
| Titration Acid (µeq/g) | <= 0.3 | 0.2 |
| Titration Base (µeq/g) | <= 0.6 | < 0.1 |
| Water (H ₂ O) | <= 0.5 % | 0.3 |
| FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL) | <= 5 | < 1 |
| ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL) | <= 10 | < 1 |

For Laboratory, Research or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: US
Packaging Site: Phillipsburg Mfg Ctr & DC



Phillipsburg, NJ 9001:2008, 14001:2004
Paris, KY 9001:2008
Mexico City, Mexico 9001:2008
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003
Gliwice, Poland 9001:2008, 17025:2005
Selangor, Malaysia 9001:2008
Dehradun, India, 9001:2008, 14001:2004, 13485:2003
Mumbai, India, 9001:2008, 17025:2005
Panoli, India 9001:2008

Richard M Siberski
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600
Avantor™ Performance Materials Inc.

3477 Corporate Parkway, Suite #200, Center Valley, PA 18034, U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610