

2-Propanol
HPLC
For use in Liquid Chromatography (HPLC & UHPLC) &
Spectrophotometry



Material No.: 9095-33
Batch No.: 0000095705
Manufactured Date: 2014/11/03
Retest Date: 2019/11/02

Certificate of Analysis

Test	Specification	Result
Ultraviolet Absorbance (1.00-cm cell vs. water) – 350 nm	<= 0.01	< 0.01
Ultraviolet Absorbance (1.00-cm cell vs. water) – 280 nm	<= 0.01	< 0.01
Ultraviolet Absorbance (1.00-cm cell vs. water) – 254 nm	<= 0.020	0.005
Ultraviolet Absorbance (1.00-cm cell vs. water) – 225 nm	<= 0.16	0.07
Ultraviolet Absorbance (1.00-cm cell vs. water) – UV Cut-off, nm	<= 205	203
Assay (CH ₃ CHOHCH ₃)	>= 99.7 %	99.9
Residue after Evaporation	<= 2.0 ppm	< 0.1
Water (by KF, coulometric)	<= 0.03 %	0.01

For Laboratory, Research or Manufacturing Use

Country of Origin: CA
Packaging Site: Paris Mfg Ctr & DC



Phillipsburg, NJ 9001:2008, 14001:2004
Paris, KY 9001:2008
Mexico City, Mexico 9001:2008
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003
Gliwice, Poland 9001:2008, 17025:2005
Selangor, Malaysia 9001:2008
Dehradun, India, 9001:2008, 14001:2004, 13485:2003
Mumbai, India, 9001:2008, 17025:2005
Panoli, India 9001:2008

Richard M Siberski
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600
Avantor™ Performance Materials Inc.

3477 Corporate Parkway, Suite #200, Center Valley, PA 18034, U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610