

Methanol  
ULTRA RESI-ANALYZED  
For Purge and Trap Analysis



Material No.: 9077-02  
Batch No.: 0000062345  
Manufactured Date: 2013/10/10  
Retest Date: 2018/10/09

## Certificate of Analysis

Test	Specification	Result
Assay (CH <sub>3</sub> OH) (by GC, corrected for water)	>= 99.9 %	100.0
Residue after Evaporation	<= 1.0000 ppm	0.2000
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.1	<0.01
Water (by KF, coulometric)	<= 0.08 %	< 0.01
Photoionization Detection (PID) Below CRQL	Passes Test	PT
Electroconductivity Detection (ELCD) Below CRQL	Passes Test	PT

For Laboratory, Research or Manufacturing Use  
Performance Tested for Use in EPA Methods  
500 Series for Drinking Water  
600 Series for Wastewater  
846 for Solid Waste

Country of Origin: US  
Packaging Site: Phillipsburg Mfg Ctr & DC



Phillipsburg, NJ 9001:2008, 14001:2004  
Paris, KY 9001:2008  
Mexico City, Mexico 9001:2008  
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003  
Gliwice, Poland 9001:2008, 17025:2005  
Selangor, Malaysia 9001:2008  
Dehradun, India, 9001:2008, 14001:2004, 13485:2003  
Mumbai, India, 9001:2008, 17025:2005  
Panoli, India 9001:2008

Richard M Siberski  
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600  
Avantor™ Performance Materials Inc.

3477 Corporate Parkway, Suite #200, Center Valley, PA 18034, U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610