

2-Propanol
HPLC
For use in Liquid Chromatography (HPLC & UHPLC) &
Spectrophotometry



Material No.: 9095-33
Batch No.: 0000111093
Manufactured Date: 2015/05/01
Retest Date: 2020/04/29

Certificate of Analysis

| Test | Specification | Result |
|--|---------------|--------|
| Ultraviolet Absorbance (1.00-cm cell vs. water) – 350 nm | <= 0.01 | < 0.01 |
| Ultraviolet Absorbance (1.00-cm cell vs. water) – 280 nm | <= 0.01 | < 0.01 |
| Ultraviolet Absorbance (1.00-cm cell vs. water) – 254 nm | <= 0.020 | 0.010 |
| Ultraviolet Absorbance (1.00-cm cell vs. water) – 225 nm | <= 0.16 | 0.14 |
| Ultraviolet Absorbance (1.00-cm cell vs. water) – UV Cut-off, nm | <= 205 | 204 |
| Assay (CH ₃ CHOHCH ₃) | >= 99.7 % | 100.0 |
| Residue after Evaporation | <= 2.0 ppm | < 0.1 |
| Water (by KF, coulometric) | <= 0.03 % | 0.01 |

For Laboratory, Research or Manufacturing Use

Country of Origin: CA
Packaging Site: Paris Mfg Ctr & DC



Phillipsburg, NJ 9001:2008, 14001:2004, FSSC 22000
Paris, KY 9001:2008
Mexico City, Mexico 9001:2008
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003
Gliwice, Poland 9001:2008, 17025:2005
Selangor, Malaysia 9001:2008
Dehradun, India, 9001:2008, 14001:2004, 13485:2003
Mumbai, India, 9001:2008, 17025:2005
Panoli, India 9001:2008

Richard M Siberski
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600
Avantor™ Performance Materials Inc.

3477 Corporate Parkway, Suite #200, Center Valley, PA 18034, U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610