

Calcium Chloride, Anhydrous
BAKER ANALYZED® A.C.S. Reagent
20 mesh and finer, for desiccant use



Material No.: 1311-05
Batch No.: 0000152115
Manufactured Date: 2015/08/31
Retest Date: 2022/08/29

Certificate of Analysis

Meets ACS Reagent Chemical Requirements,

Test	Specification	Result
Assay (CaCl ₂) (by Ag titrn)	>= 96.0 %	98.8
Titration Base (meq/g)	<= 0.006	0.004
Mesh - On U.S. No. 20 Sieve	<= 10 %	< 1
Mesh - Thru U.S. No. 200 Sieve	<= 30 %	8

For Laboratory, Research or Manufacturing Use
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: JP
Packaging Site: Paris Mfg Ctr & DC



Phillipsburg, NJ 9001:2008, 14001:2004, FSSC 22000
Paris, KY 9001:2008
Mexico City, Mexico 9001:2008
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003
Gliwice, Poland 9001:2008, 13485:2012
Selangor, Malaysia 9001:2008
Dehradun, India, 9001:2008, 14001:2004, 13485:2003
Mumbai, India, 9001:2008
Panoli, India 9001:2008


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600
Avantor™ Performance Materials Inc.

3477 Corporate Parkway, Suite #200, Center Valley, PA 18034, U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610