

Acetonitrile
HPLC
For use in Liquid Chromatography and Spectrophotometry



Material No.: 9012-03
Batch No.: 0000159420
Manufactured Date: 2016/11/16
Retest Date: 2021/11/15

Certificate of Analysis

Test	Specification	Result
Appearance	Passes Test	PT
Assay (GC)	$\geq 99.9\%$	100.0
Color (APHA)	≤ 10	5
Water (H ₂ O)	$\leq 0.02\%$	< 0.01
Gradient Elution Test (a.u.) – 254 nm	≤ 0.005	0.002
Residue after Evaporation	≤ 3 ppm	< 1
Titration Acid ($\mu\text{eq/g}$)	≤ 8.0	< 1.0
Titration Base ($\mu\text{eq/g}$)	≤ 0.6	< 0.1
Ultraviolet Absorbance (1.00 cm) – 200 nm	≤ 0.1	< 0.1
Ultraviolet Absorbance (1.00 cm) – 210 nm	≤ 0.05	0.02
Ultraviolet Absorbance (1.00 cm) – 220 nm	≤ 0.03	0.01
Ultraviolet Absorbance (1.00 cm) – 254 nm	≤ 0.01	< 0.01
Ultraviolet Absorbance (1.00 cm) – 280 nm	≤ 0.01	< 0.01
Ultraviolet Absorbance (1.00 cm) – 350 nm	≤ 0.01	< 0.01
Ultraviolet Absorbance (1.00 cm) – 400 nm	≤ 0.01	< 0.01
Ultraviolet Absorbance (1.00 cm) – UV Cut-off, nm	≤ 190	188

For Laboratory, Research or Manufacturing Use
Filtered through a 0.2 micron filter.

Country of Origin: US
Packaging Site: Paris Mfg Ctr & DC



Phillipsburg, NJ 9001:2008, 14001:2004, FSSC 22000
Paris, KY 9001:2008
Mexico City, Mexico 9001:2008
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003
Gliwice, Poland 9001:2008, 13485:2012
Selangor, Malaysia 9001:2008
Dehradun, India, 9001:2008, 14001:2004, 13485:2003
Mumbai, India, 9001:2008
Panoli, India 9001:2008

Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600
Avantor™ Performance Materials Inc.

3477 Corporate Parkway, Suite #200, Center Valley, PA 18034, U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610