

Acetonitrile  
HPLC  
For use in Liquid Chromatography and Spectrophotometry



Material No.: 9012-03  
Batch No.: 0000166378  
Manufactured Date: 2017/02/17  
Retest Date: 2022/02/16

## Certificate of Analysis

Test	Specification	Result
Appearance	Passes Test	PT
Assay (GC)	$\geq 99.9\%$	100.0
Color (APHA)	$\leq 10$	5
Water (H <sub>2</sub> O)	$\leq 0.02\%$	$< 0.01$
Gradient Elution Test (a.u.) – 254 nm	$\leq 0.005$	0.002
Residue after Evaporation	$\leq 3$ ppm	$< 1$
Titration Acid ( $\mu\text{eq/g}$ )	$\leq 8.0$	$< 1.0$
Titration Base ( $\mu\text{eq/g}$ )	$\leq 0.6$	$< 0.1$
Ultraviolet Absorbance (1.00 cm) – 200 nm	$\leq 0.1$	$< 0.1$
Ultraviolet Absorbance (1.00 cm) – 210 nm	$\leq 0.05$	0.01
Ultraviolet Absorbance (1.00 cm) – 220 nm	$\leq 0.03$	$< 0.01$
Ultraviolet Absorbance (1.00 cm) – 254 nm	$\leq 0.01$	$< 0.01$
Ultraviolet Absorbance (1.00 cm) – 280 nm	$\leq 0.01$	$< 0.01$
Ultraviolet Absorbance (1.00 cm) – 350 nm	$\leq 0.01$	$< 0.01$
Ultraviolet Absorbance (1.00 cm) – 400 nm	$\leq 0.01$	$< 0.01$
Ultraviolet Absorbance (1.00 cm) – UV Cut-off, nm	$\leq 190$	187

For Laboratory, Research or Manufacturing Use  
Filtered through a 0.2 micron filter.

Country of Origin: US  
Packaging Site: Paris Mfg Ctr & DC



Phillipsburg, NJ 9001:2008, 14001:2004, FSSC 22000  
Paris, KY 9001:2008  
Mexico City, Mexico 9001:2008  
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003  
Gliwice, Poland 9001:2008, 13485:2012  
Selangor, Malaysia 9001:2008  
Dehradun, India, 9001:2008, 14001:2004, 13485:2003  
Mumbai, India, 9001:2008  
Panoli, India 9001:2008

Jamie Ethier  
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600  
Avantor Performance Materials, LLC.

3477 Corporate Parkway. Center Valley, PA 18034. U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610