

Acetone
HPLC
For use in Liquid Chromatography (HPLC & UHPLC) &
Spectrophotometry



Material No.: 9002-02
Batch No.: 0000189588
Manufactured Date: 2017/11/20
Retest Date: 2022/11/19

Certificate of Analysis

Test	Specification	Result
Ultraviolet Absorbance (1.00-cm cell vs. water) – 400–350 nm	<= 0.01	< 0.01
Ultraviolet Absorbance (1.00-cm cell vs. water) – UV Cut-off, nm	<= 330	328
Fluorescence Trace Impurities, measured as Quinine Base – at 450 nm Emission	<= 0.3 ppb	0.1
Fluorescence Trace Impurities, measured as Quinine Base – at Emission Maximum for Impurities	<= 1.0 ppb	0.1
Assay ((CH ₃) ₂ CO)(by GC)	>= 99.3 %	100.0
Residue after Evaporation	<= 2.0 ppm	< 2.0
Titration Acid (µeq/g)	<= 0.3	< 0.3
Titration Base (µeq/g)	<= 0.6	< 0.6
Water (H ₂ O)	<= 0.4 %	0.3

For Laboratory, Research or Manufacturing Use
Filtered through a 0.2 micron filter.
Packaged under Nitrogen.

Country of Origin: US
Packaging Site: Paris Mfg Ctr & DC



Phillipsburg, NJ 9001:2008, 14001:2004, FSSC 22000
Paris, KY 9001:2008
Mexico City, Mexico 9001:2008
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003
Gliwice, Poland 9001:2008, 13485:2012
Selangor, Malaysia 9001:2008
Dehradun, India, 9001:2008, 14001:2004, 13485:2003
Mumbai, India, 9001:2008
Panoli, India 9001:2008

Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600
Avantor Performance Materials, LLC.

3477 Corporate Parkway. Center Valley, PA 18034. U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610