Acetonitrile BAKER ANALYZED® HPLC Ultra Gradient Solvent

For use in Liquid Chromatography (HPLC & UHPLC) & Spectrophotometry



Material No.: 9017-33 Batch No.: 0000195204

Manufactured Date: 2018/02/24 Retest Date: 2023/02/23

Revision No: 1

Certificate of Analysis

Test	Specification	Result
ltraviolet Absorbance (1.00-cm cell vs. water) - 00-254 nm	<= 0.005	0.002
ltraviolet Absorbance (1.00-cm cell vs. water) - 220 m	<= 0.01	0.01
traviolet Absorbance (1.00-cm cell vs. water) - 200 n	<= 0.05	0.02
traviolet Absorbance (1.00-cm cell vs. water) - UV ut-off, nm	<= 190	188
radient Elution Test (a.u.) – 254 nm	<= 0.0005	< 0.0005
radient Elution Test (a.u.) – 210 nm	<= 0.002	< 0.002
ensity (g/mL) at 25°C	0.775 - 0.780	0.779
uorescence Trace Impurities, measured as Quinine ase – at 450 nm Emission	<= 0.3 ppb	0.1
uorescence Trace Impurities, measured as Quinine use – at Emission Maximum for Impurities	<= 1.0 ppb	0.1
ssay (CH3CN) (by GC)	>= 99.9 %	100.0
ppearance	Passes Test	РТ
olor (APHA)	<= 10	5
uorescence Trace Impurities, measured as Quinine ase – Fluorescence detection (PAH)	<= 0.5 ppb	0.1
esidue after Evaporation	<= 1.0 ppm	< 0.1
trable Acid (µeq/g)	<= 0.8	0.8
rrable Base (µeq/g)	<= 0.6	0.6
ater (by KF, coulometric)	<= 100 ppm	11
arbonyl Compounds (as Acetone)	<= 25 ppb	20

For Laboratory, Research or Manufacturing Use

Material No.: 9017-33 Batch No.: 0000195204

Filtered through a 0.2 micron filter.

Country of Origin: US

Packaging Site: Paris Mfg Ctr & DC



Phillipsburg, NJ 9001:2008, 14001:2004, FSSC 22000
Paris, KY 9001:2008
Mexico City, Mexico 9001:2008
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003
Gliwice, Poland 9001:2008, 13485:2012
Selangor, Malaysia 9001:2008
Dehradun, India, 9001:2008, 14001:2004, 13485:2003
Mumbai, India, 9001:2008
Panoli, India 9001:2008

Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600 Avantor Performance Materials, LLC.