

2-Propanol  
HPLC  
For use in Liquid Chromatography (HPLC & UHPLC) &  
Spectrophotometry



Material No.: 9095-33  
Batch No.: 0000192841  
Manufactured Date: 2018/01/22  
Retest Date: 2023/01/21  
Revision No: 1

## Certificate of Analysis

Test	Specification	Result
Ultraviolet Absorbance (1.00-cm cell vs. water) – 350 nm	<= 0.01	< 0.01
Ultraviolet Absorbance (1.00-cm cell vs. water) – 280 nm	<= 0.01	< 0.01
Ultraviolet Absorbance (1.00-cm cell vs. water) – 254 nm	<= 0.020	0.010
Ultraviolet Absorbance (1.00-cm cell vs. water) – 225 nm	<= 0.16	0.12
Ultraviolet Absorbance (1.00-cm cell vs. water) – UV Cut-off, nm	<= 205	205
Assay (CH <sub>3</sub> CHOHCH <sub>3</sub> )	>= 99.7 %	100.0
Residue after Evaporation	<= 2.0 ppm	< 1.0
Water (by KF, coulometric)	<= 0.03 %	< 0.01

For Laboratory, Research or Manufacturing Use

Country of Origin: CA  
Packaging Site: Paris Mfg Ctr & DC



Phillipsburg, NJ 9001:2008, 14001:2004, FSSC 22000  
Paris, KY 9001:2008  
Mexico City, Mexico 9001:2008  
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003  
Gliwice, Poland 9001:2008, 13485:2012  
Selangor, Malaysia 9001:2008  
Dehradun, India, 9001:2008, 14001:2004, 13485:2003  
Mumbai, India, 9001:2008  
Panoli, India 9001:2008

  
Jamie Ethier  
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600  
Avantor Performance Materials, LLC.

3477 Corporate Parkway. Center Valley, PA 18034. U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610