

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 0000196626
Manufactured Date: 2018/04/04
Expiration Date: 2020/04/01
Revision No: 1

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	>= 99.9 %	100.0
Residue after Evaporation	<= 1.0000 ppm	< 0.1000
Titration Acid (µeq/g)	<= 0.3	0.1
Titration Base (µeq/g)	<= 0.1	<0.01
Water (by KF, coulometric)	<= 0.08 %	< 0.01
Photoionization Detection (PID) Below CRQL	Passes Test	PT
Electroconductivity Detection (ELCD) Below CRQL	Passes Test	PT

For Laboratory, Research or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: US
Packaging Site: Phillipsburg Mfg Ctr & DC



Phillipsburg, NJ 9001:2008, 14001:2004, FSSC 22000
Paris, KY 9001:2008
Mexico City, Mexico 9001:2008
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003
Gliwice, Poland 9001:2008, 13485:2012
Selangor, Malaysia 9001:2008
Dehradun, India, 9001:2008, 14001:2004, 13485:2003
Mumbai, India, 9001:2008
Panoli, India 9001:2008


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600
Avantor Performance Materials, LLC.

3477 Corporate Parkway. Center Valley, PA 18034. U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610