

Acetone  
HPLC  
For use in Liquid Chromatography (HPLC & UHPLC) &  
Spectrophotometry



Material No.: 9002-33  
Batch No.: 0000201838  
Manufactured Date: 2018/05/29  
Retest Date: 2023/05/28  
Revision No: 1

## Certificate of Analysis

Test	Specification	Result
Ultraviolet Absorbance (1.00-cm cell vs. water) – 400–350 nm	<= 0.01	< 0.01
Ultraviolet Absorbance (1.00-cm cell vs. water) – UV Cut-off, nm	<= 330	328
Fluorescence Trace Impurities, measured as Quinine Base – at 450 nm Emission	<= 0.3 ppb	0.1
Fluorescence Trace Impurities, measured as Quinine Base – at Emission Maximum for Impurities	<= 1.0 ppb	0.1
Assay ((CH <sub>3</sub> ) <sub>2</sub> CO)(by GC)	>= 99.3 %	100.0
Residue after Evaporation	<= 2.0 ppm	< 1.0
Titration Acid (µeq/g)	<= 0.3	< 0.1
Titration Base (µeq/g)	<= 0.6	0.1
Water (H <sub>2</sub> O)	<= 0.4 %	0.1

For Laboratory, Research or Manufacturing Use  
Filtered through a 0.2 micron filter.  
Packaged under Nitrogen.

Country of Origin: US  
Packaging Site: Paris Mfg Ctr & DC



Phillipsburg, NJ 9001:2008, 14001:2004, FSSC 22000  
Paris, KY 9001:2008  
Mexico City, Mexico 9001:2008  
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003  
Gliwice, Poland 9001:2008, 13485:2012  
Selangor, Malaysia 9001:2008  
Dehradun, India, 9001:2008, 14001:2004, 13485:2003  
Mumbai, India, 9001:2008  
Panoli, India 9001:2008

*James Ethier*  
Jamie Ethier  
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600  
Avantor Performance Materials, LLC.

3477 Corporate Parkway. Center Valley, PA 18034. U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610