

2-Propanol
HPLC
For use in Liquid Chromatography (HPLC & UHPLC) &
Spectrophotometry



Material No.: 9095-33
Batch No.: 0000207365
Manufactured Date: 2018/08/02
Retest Date: 2023/08/01
Revision No: 1

Certificate of Analysis

Test	Specification	Result
Ultraviolet Absorbance (1.00-cm cell vs. water) – 350 nm	<= 0.01	< 0.01
Ultraviolet Absorbance (1.00-cm cell vs. water) – 280 nm	<= 0.01	< 0.01
Ultraviolet Absorbance (1.00-cm cell vs. water) – 254 nm	<= 0.020	0.011
Ultraviolet Absorbance (1.00-cm cell vs. water) – 225 nm	<= 0.16	0.13
Ultraviolet Absorbance (1.00-cm cell vs. water) – UV Cut-off, nm	<= 205	204
Assay (CH ₃ CHOHCH ₃)	>= 99.7 %	100.0
Residue after Evaporation	<= 2.0 ppm	< 1.0
Water (by KF, coulometric)	<= 0.03 %	0.03

For Laboratory, Research or Manufacturing Use

Country of Origin: CA
Packaging Site: Paris Mfg Ctr & DC



Phillipsburg, NJ 9001:2015, FSSC 22000
Paris, KY 9001:2008
Mexico City, Mexico 9001:2008
Gliwice, Poland 9001:2008
Selangor, Malaysia 9001:2008
Dehradun, India 9001:2015, 14001:2015, 13485:2016
Mumbai, India 9001:2015
Panoli, India 9001:2015

Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600
Avantor Performance Materials, LLC.

3477 Corporate Parkway. Center Valley, PA 18034. U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610