

1,2-Dimethoxyethane
BAKER ANALYZED® Reagent

(ethylene glycol dimethyl ether)



Material No.: J331-07
Batch No.: 0000026798
Manufactured Date: 2011/07/28
Retest Date: 2016/07/26

Certificate of Analysis

Test	Specification	Result
Assay (C ₄ H ₁₀ O ₂) (by GC, corrected for water)	>= 99.0 %	99.8
Acidity (as CH ₃ COOH)	<= 0.015 %	0.002
Density (g/mL) at 20°C	0.865 – 0.871	0.866
Identification (by IR)	Passes Test	PT
Peroxide (as H ₂ O ₂)	<= 25 ppm	21
Refractive Index, n _D ²⁰	1.379 – 1.381	1.380
Residue after Evaporation	<= 0.0020 %	< 0.0010
Water (H ₂ O)(by Karl Fischer titrn)	<= 0.2 %	0.2
Boiling Range (1–95 mL)Lower	83 – 86 °C	85
Boiling Range (1–95 mL)Upper	83 – 86 °C	86

For Laboratory, Research or Manufacturing Use

Country of Origin: US
Packaging Site: Paris Mfg Ctr & DC



Phillipsburg, NJ 9001:2008, 14001:2004
Paris, KY 9001:2008
Mexico City, Mexico 9001:2008
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003
Gliwice, Poland 9001:2008, 17025:2005
Selangor, Malaysia 9001:2008
Dehradun, India, 9001:2008, 14001:2004, 13485:2003
Mumbai, India, 9001:2008, 17025:2005
Panoli, India 9001:2008

Richard M Siberski
Global Director of Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600
Avantor™ Performance Materials Inc.

3477 Corporate Parkway, Suite #200, Center Valley, PA 18034, U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610