

Ammonium Acetate  
HPLC  
For Use in High Performance Liquid Chromatography



Material No.: 0599-08  
Batch No.: 0000239123  
Manufactured Date: 2018/08/06  
Retest Date: 2025/08/04  
Revision No: 1

## Certificate of Analysis

Test	Specification	Result
Assay (CH <sub>3</sub> COONH <sub>4</sub> ) (by acid-base titration)	>= 97 %	100
ACS - Insoluble Matter	<= 0.005 %	0.001
Residue after Ignition	<= 0.010 %	0.003
pH of 5% Solution at 25°C	6.7 - 7.3	7.0
Nitrate (NO <sub>3</sub> )	<= 0.001 %	< 0.001
Sulfate (SO <sub>4</sub> )	<= 0.001 %	< 0.001
UV Absorbance (1 M Aqueous Solution) (1.00-cm cell vs. water) - 350 nm	<= 0.010	0.003
UV Absorbance (1 M Aqueous Solution) (1.00-cm cell vs. water) - 280 nm	<= 0.010	0.007
UV Absorbance (1 M Aqueous Solution) (1.00-cm cell vs. water) - 254 nm	<= 0.020	0.012
Chloride (Cl)	<= 5 ppm	< 5
Trace Impurities - ACS - Heavy Metals (as Pb)	<= 5 ppm	< 5
Trace Impurities - Iron (Fe)	<= 5 ppm	< 5

For Laboratory, Research or Manufacturing Use

Country of Origin: MX  
Packaging Site: Paris Mfg Ctr & DC

  
Jamie Ethier  
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700  
Avantor Performance Materials, LLC  
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700