

Calcium Chloride, Anhydrous
BAKER ANALYZED® A.C.S. Reagent
20 mesh and finer, for desiccant use



Material No.: 1311-05
Batch No.: 0000259291
Manufactured Date: 2020/01/06
Retest Date: 2027/01/04
Revision No: 1

Certificate of Analysis

Meets ACS Reagent Chemical Requirements,

Test	Specification	Result
Assay (CaCl ₂) (by Ag titrn)	$\geq 96.0 \%$	100.4
Titration Base (meq/g)	≤ 0.006	0.001
Mesh - On U.S. No. 20 Sieve	$\leq 10 \%$	1
Mesh - Thru U.S. No. 200 Sieve	$\leq 30 \%$	9

For Laboratory, Research or Manufacturing Use
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: JP
Packaging Site: Paris Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700