

Acetonitrile
BAKER ANALYZED® HPLC Ultra Gradient Solvent
For use in Liquid Chromatography (HPLC & UHPLC) &
Spectrophotometry



Material No.: 9017-33
Batch No.: 0000275980
Manufactured Date: 2020/11/05
Retest Date: 2025/11/04
Revision No: 1

Certificate of Analysis

Test	Specification	Result
Ultraviolet Absorbance (1.00-cm cell vs. water) – 400–254 nm	<= 0.005	0.004
Ultraviolet Absorbance (1.00-cm cell vs. water) – 220 nm	<= 0.01	0.01
Ultraviolet Absorbance (1.00-cm cell vs. water) – 200 nm	<= 0.05	0.02
Ultraviolet Absorbance (1.00-cm cell vs. water) – UV Cut-off, nm	<= 190	188
Gradient Elution Test (a.u.) – 254 nm	<= 0.0005	< 0.0001
Gradient Elution Test (a.u.) – 210 nm	<= 0.002	< 0.001
Density (g/mL) at 25°C	0.775 – 0.780	0.777
Fluorescence Trace Impurities, measured as Quinine Base – at 450 nm Emission	<= 0.3 ppb	0.1
Fluorescence Trace Impurities, measured as Quinine Base – at Emission Maximum for Impurities	<= 1.0 ppb	0.6
Assay (CH ₃ CN) (by GC)	>= 99.9 %	100.0
Appearance	Passes Test	PT
Color (APHA)	<= 10	5
Fluorescence Trace Impurities, measured as Quinine Base – Fluorescence detection (PAH)	<= 0.5 ppb	0.3
Residue after Evaporation	<= 1.0 ppm	0.4
Titration Acid (µeq/g)	<= 0.8	0.2
Titration Base (µeq/g)	<= 0.6	< 0.1
Water (by KF, coulometric)	<= 100 ppm	< 10
Carbonyl Compounds (as Acetone)	<= 25 ppb	20

For Laboratory, Research or Manufacturing Use

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Material No.: 9017-33
Batch No.: 0000275980

Filtered through a 0.2 micron filter.

Country of Origin: US
Packaging Site: Phillipsburg Mfg Ctr & DC



Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700