

Methanol
HPLC
For use in Liquid Chromatography (HPLC & UHPLC) &
Spectrophotometry
(methyl alcohol)



Material No.: 9093-68
Batch No.: 0000243087
Manufactured Date: 2019/09/24
Expiration Date: 2022/09/23
Revision No: 1

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	>= 99.9 %	100.0
Ultraviolet Absorbance (1.00-cm cell vs. water) - 400-254 nm	<= 0.01	0.01
Ultraviolet Absorbance (1.00-cm cell vs. water) - 225 nm	<= 0.15	0.14
Ultraviolet Absorbance (1.00-cm cell vs. water) - UV Cut-off, nm	<= 205	204
Gradient Elution Test (a.u.) - 254 nm	<= 0.002	0.001
Fluorescence Trace Impurities, measured as Quinine Base - at 450 nm Emission	<= 0.3 ppb	0.1
Fluorescence Trace Impurities, measured as Quinine Base - at Emission Maximum for Impurities	<= 1.0 ppb	0.4
Acetone	<= 0.001 %	< 0.001
Residue after Evaporation	<= 1.0000 ppm	0.3000
Titration Acid (µeq/g)	<= 0.3	0.3
Titration Base (µeq/g)	<= 0.1	<0.01
Water (by KF, coulometric)	<= 0.05 %	< 0.01

For Laboratory, Research or Manufacturing Use
Filtered through a 0.2 micron filter.

Country of Origin: US
Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700