

Hydrogen Peroxide, 30%
BAKER ANALYZED® A.C.S. Reagent
(Stabilized)



Material No.: 2186-03
Batch No.: 25C0461082
Manufactured Date: 2025-02-24
Expiration Date: 2026-08-25
Revision No.: 0

Certificate of Analysis

Meets ACS Reagent Chemical Requirements

Test	Specification	Result
Assay (H ₂ O ₂)	30.0 – 32.0 %	31.6 %
Color (APHA)	<= 10	<5
Residue after Evaporation	<= 0.0020 %	0.0002 %
Titration Acid (meq/g)	<= 0.0006	<0.0001
Chloride (Cl)	<= 2 ppm	<0 ppm
Nitrate (NO ₃)	<= 2 ppm	<2 ppm
Phosphate (PO ₄)	<= 2 ppm	1 ppm
ACS – Sulfate (SO ₄)	<= 5 ppm	<3 ppm
Ammonium (NH ₄)	<= 5 ppm	<3 ppm
Trace Impurities – Copper (Cu)	<= 50 ppb	<1 ppb
Trace Impurities – Iron (Fe)	<= 100.0 ppb	2.6 ppb
Heavy Metals (as Pb)	<= 1000.0 ppb	<10.0 ppb
Trace Impurities – Lead (Pb)	<= 100 ppb	<10 ppb
Trace Impurities – Nickel (Ni)	<= 50 ppb	<5 ppb

For Laboratory, Research, or Manufacturing Use
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: United States
Packaging Site: Paris Mfg Ctr & DC


Michelle Bales
Sr. Manager, Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700