

Material No.: 2818-05  
Batch No.: 23E1961036  
Manufactured Date: 2022-11-01  
Retest Date: 2029-10-30  
Revision No.: 0

## Certificate of Analysis

Test	Specification	Result
Assay (as C <sub>8</sub> H <sub>17</sub> SO <sub>3</sub> Na) (by acidimetry)	≥ 98.0 %	99.9 %
UV Absorbance of 0.25 M Solution - 250 nm	≤ 0.050	0.007
UV Absorbance of 0.25 M Solution - 240 nm	≤ 0.050	0.008
UV Absorbance of 0.25 M Solution - 230 nm	≤ 0.050	0.014
UV Absorbance of 0.25 M Solution - 220 nm	≤ 0.060	0.026
UV Absorbance of 0.25 M Solution - 210 nm	≤ 0.080	0.059
UV Absorbance of 0.25 M Solution - 200 nm	≤ 0.200	0.135

For Laboratory, Research, or Manufacturing Use  
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: United Kingdom  
Packaging Site: Paris Mfg Ctr & DC



Jamie Ethier  
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700

Page 1 of 1