

Material No.: 2818-05  
Batch No.: 23J0961038  
Manufactured Date: 2023-02-01  
Retest Date: 2030-01-30  
Revision No.: 0

## Certificate of Analysis

Test	Specification	Result
Assay (as $C_8H_{17}SO_3Na$ ) (by acidimetry)	$\geq 98.0\%$	99.5%
UV Absorbance of 0.25 M Solution - 250 nm	$\leq 0.050$	0.009
UV Absorbance of 0.25 M Solution - 240 nm	$\leq 0.050$	0.009
UV Absorbance of 0.25 M Solution - 230 nm	$\leq 0.050$	0.013
UV Absorbance of 0.25 M Solution - 220 nm	$\leq 0.060$	0.019
UV Absorbance of 0.25 M Solution - 210 nm	$\leq 0.080$	0.035
UV Absorbance of 0.25 M Solution - 200 nm	$\leq 0.200$	0.090

For Laboratory, Research, or Manufacturing Use  
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: United Kingdom  
Packaging Site: Paris Mfg Ctr & DC



Michelle Bales  
Sr. Manager, Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700