

Material No.: 2818-05
Batch No.: 23J2661015
Manufactured Date: 2023-02-01
Retest Date: 2030-01-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (as C ₈ H ₁₇ SO ₃ Na) (by acidimetry)	≥ 98.0 %	99.5 %
UV Absorbance of 0.25 M Solution - 250 nm	≤ 0.050	0.009
UV Absorbance of 0.25 M Solution - 240 nm	≤ 0.050	0.009
UV Absorbance of 0.25 M Solution - 230 nm	≤ 0.050	0.013
UV Absorbance of 0.25 M Solution - 220 nm	≤ 0.060	0.019
UV Absorbance of 0.25 M Solution - 210 nm	≤ 0.080	0.035
UV Absorbance of 0.25 M Solution - 200 nm	≤ 0.200	0.090

For Laboratory, Research, or Manufacturing Use
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: United Kingdom
Packaging Site: Paris Mfg Ctr & DC



Michelle Bales
Sr. Manager, Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700