

Material No.: 2818-05  
Batch No.: 24B1961005  
Manufactured Date: 2023-08-01  
Retest Date: 2030-07-30  
Revision No.: 0

## Certificate of Analysis

Test	Specification	Result
Assay (as C <sub>8</sub> H <sub>17</sub> SO <sub>3</sub> Na) (by acidimetry)	≥ 98.0 %	100.1 %
UV Absorbance of 0.25 M Solution - 250 nm	≤ 0.050	0.008
UV Absorbance of 0.25 M Solution - 240 nm	≤ 0.050	0.010
UV Absorbance of 0.25 M Solution - 230 nm	≤ 0.050	0.018
UV Absorbance of 0.25 M Solution - 220 nm	≤ 0.060	0.030
UV Absorbance of 0.25 M Solution - 210 nm	≤ 0.080	0.051
UV Absorbance of 0.25 M Solution - 200 nm	≤ 0.200	0.127

For Laboratory, Research, or Manufacturing Use  
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: United Kingdom  
Packaging Site: Paris Mfg Ctr & DC



Michelle Bales  
Sr. Manager, Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700