

Acetone
HPLC
For use in Liquid Chromatography (HPLC & UHPLC) & Spectrophotometry



Material No.: 9002-33
Batch No.: 22A1262004
Manufactured Date: 2021-12-15
Retest Date: 2026-12-14
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Ultraviolet Absorbance (1.00-cm cell vs. water) - 400-350 nm	≤ 0.01	< 0.01
Ultraviolet Absorbance (1.00-cm cell vs. water) - UV Cut-off, nm	≤ 330	329
Fluorescence Trace Impurities, measured as Quinine Base - at 450 nm Emission	≤ 0.3 ppb	0.1 ppb
Fluorescence Trace Impurities, measured as Quinine Base - at Emission Maximum for Impurities	≤ 1.0 ppb	0.2 ppb
Assay ((CH ₃) ₂ CO)(by GC)	≥ 99.3 %	100.0 %
Residue after Evaporation	≤ 2.0 ppm	0.1 ppm
Titration Acid (μeq/g)	≤ 0.3	0.3
Titration Base (μeq/g)	≤ 0.6	0.1
Water (H ₂ O)	≤ 0.4 %	0.3 %

For Laboratory, Research, or Manufacturing Use
Filtered through a 0.2 micron filter.
Packaged under Nitrogen.
Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700