

Acetonitrile  
BAKER HPLC ANALYZED  
For use in Liquid Chromatography and  
Spectrophotometry



Material No.: 9012-03  
Batch No.: 23D2661004  
Manufactured Date: 2023-03-30  
Retest Date: 2028-03-28  
Revision No.: 0

## Certificate of Analysis

Test	Specification	Result
Appearance	Passes Test	Passes Test
Assay (GC)	≥ 99.9 %	100.0 %
Color (APHA)	≤ 10	< 5
Water (H <sub>2</sub> O)	≤ 0.02 %	< 0.01 %
Gradient Elution Test (a.u.) – 254 nm	≤ 0.005	< 0.005
Residue after Evaporation	≤ 3 ppm	< 1 ppm
Titration Acid (μeq/g)	≤ 8	< 1
Titration Base (μeq/g)	≤ 0.6	< 0.1
Ultraviolet Absorbance (1.00 cm) – 200 nm	≤ 0.1	< 0.1
Ultraviolet Absorbance (1.00 cm) – 210 nm	≤ 0.05	< 0.01
Ultraviolet Absorbance (1.00 cm) – 220 nm	≤ 0.03	< 0.01
Ultraviolet Absorbance (1.00 cm) – 254 nm	≤ 0.01	< 0.01
Ultraviolet Absorbance (1.00 cm) – 280 nm	≤ 0.01	< 0.01
Ultraviolet Absorbance (1.00 cm) – 350 nm	≤ 0.01	< 0.01
Ultraviolet Absorbance (1.00 cm) – 400 nm	≤ 0.01	< 0.01
Ultraviolet Absorbance (1.00 cm) – UV Cut-off, nm	≤ 190	< 190

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC  
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700

Acetonitrile  
BAKER HPLC ANALYZED  
For use in Liquid Chromatography and  
Spectrophotometry



Material No.: 9012-03  
Batch No.: 23D2661004

---

Test	Specification	Result
------	---------------	--------

---

---

For Laboratory, Research, or Manufacturing Use  
Filtered through a 0.2 micron filter.

Country of Origin: USA  
Packaging Site: Paris Mfg Ctr & DC

Jamie Ethier  
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700  
Avantor Performance Materials, LLC  
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700