

Methanol
HPLC
Gradient Grade: For use in Liquid Chromatography
(HPLC & UHPLC) & Spectrophotometry
(methyl alcohol)



Material No.: 9093-68
Batch No.: 21K1062025
Manufactured Date: 2021-10-27
Expiration Date: 2024-10-26
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Ultraviolet Absorbance (1.00-cm cell vs. water) - 400-254 nm	≤ 0.01	0.01
Ultraviolet Absorbance (1.00-cm cell vs. water) - 225 nm	≤ 0.15	0.15
Ultraviolet Absorbance (1.00-cm cell vs. water) - UV Cut-off, nm	≤ 205	204
Gradient Elution Test (a.u.) - 254 nm	≤ 0.002	< 0.001
Fluorescence Trace Impurities, measured as Quinine Base - at 450 nm Emission	≤ 0.3 ppb	< 0.1 ppb
Fluorescence Trace Impurities, measured as Quinine Base - at Emission Maximum for Impurities	≤ 1.0 ppb	0.3 ppb
Acetone	≤ 0.001 %	< 0.001 %
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Titration Acid (μeq/g)	≤ 0.3	0.3
Titration Base (μeq/g)	≤ 0.10	< 0.02
Water (by KF, coulometric)	≤ 0.05 %	< 0.01 %

For Laboratory, Research, or Manufacturing Use
Filtered through a 0.2 micron filter.
Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700