Methanol HPLC Gradient Grade: For use in Liquid Chromatography (HPLC & UHPLC) & Spectrophotometry (methyl alcohol)





Material No.: 9093-68 Batch No.: 22E0762002 Manufactured Date: 2022-04-20 Expiration Date: 2025-04-19 Revision No.: 0

## Certificate of Analysis

| Test  | Specification   | Result    |
|---|-----------------|-----------|
| Assay (CH3OH) (by GC, corrected for water)  | ≥ <b>99.9</b> % | 100.0 %   |
| Ultraviolet Absorbance (1.00-cm cell vs. water) - 400-254 nm  | ≤ <b>0.01</b>   | 0.01      |
| Ultraviolet Absorbance (1.00-cm cell vs. water) - 225 nm  | ≤ 0.15          | 0.13      |
| Ultraviolet Absorbance (1.00-cm cell vs. water) - UV Cut-off, nm  | ≤ 205           | 204       |
| Gradient Elution Test (a.u.) – 254 nm   | ≤ 0.002         | < 0.001   |
| Fluorescence Trace Impurities, measuredFluorescence Trace Impurities, measured as Quinine Base – at 450 nm Emission                 | $\leq$ 0.3 ppb  | 0.3 ppb   |
| Fluorescence Trace Impurities, measuredFluorescence Trace Impurities, measured as Quinine Base – at Emission Maximum for Impurities | $\leq$ 1.0 ppb  | 0.3 ppb   |
| Acetone   | $\leq$ 0.001 %  | < 0.001 % |
| Residue after Evaporation   | $\leq$ 1.0 ppm  | 0.7 ppm   |
| Titrable Acid (µeq/g)   | ≤ 0.3           | 0.3       |
| Titrable Base (µeq/g)   | ≤ <b>0.10</b>   | < 0.02    |
| Water (by KF, coulometric)  | $\leq$ 0.05 %   | < 0.01 %  |
|   |                 |           |

For Laboratory,Research,or Manufacturing Use Filtered through a 0.2 micron filter.

Country of Origin: USA Packaging Site: Phillipsburg Mfg Ctr & DC

ames Techie

Jamie Ethier Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700 Avantor Performance Materials, LLC 100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700 Page 1 of 1