

Methanol
HPLC
Gradient Grade: For use in Liquid Chromatography
(HPLC & UHPLC) & Spectrophotometry
(methyl alcohol)



Material No.: 9093-68
Batch No.: T22L100008
Manufactured Date: 2022-11-28
Expiration Date: 2025-11-27
Revision No.: 0

Certificate of Analysis

| Test | Specification | Result |
|---|---------------|-----------|
| Assay (CH ₃ OH) (by GC, corrected for water) | ≥ 99.9 % | 100.0 % |
| Ultraviolet Absorbance (1.00-cm cell v) | ≤ 0.01 | < 0.01 |
| Ultraviolet Absorbance (1.00-cm cell v) | ≤ 0.15 | ≤ 0.15 |
| Ultraviolet Absorbance (1.00-cm cell v) | ≤ 205 | 205 |
| Gradient Elution Test (a.u.) - 254 nm | ≤ 0.002 | ≤ 0.002 |
| Fluorescence Trace Impurities, measured | ≤ 0.3 ppb | 0.1 ppb |
| Fluorescence Trace Impurities, measured | ≤ 1.0 ppb | 0.5 ppb |
| Acetone | ≤ 0.001 % | ≤ 0.001 % |
| Residue after Evaporation | ≤ 1.0 ppm | 0.4 ppm |
| Titration Acid (µeq/g) | ≤ 0.30 | 0.20 |
| Titration Base (µeq/g) | ≤ 0.10 | 0.05 |
| Water (by KF, coulometric) | ≤ 0.05 % | 0.04 % |

For Laboratory, Research, or Manufacturing Use
Filtered through a 0.2 micron filter.

Country of Origin: India


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700

Page 1 of 1