

Material No.: 9515-03
Batch No.: 23J0262009
Manufactured Date: 2023-09-05
Retest Date: 2028-09-03
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (by GC, corrected for water)	$\geq 99.7 \%$	99.9 %
Ultraviolet Absorbance (1.00-cm cell vs. water) - 280 nm	≤ 0.05	0.03
Ultraviolet Absorbance (1.00-cm cell vs. water) - 350 nm	≤ 0.01	0.01
Residue after Evaporation	$\leq 0.0005 \%$	0.0002 %
Water (by Karl Fischer titrn)	$\leq 0.1 \%$	< 0.1 %

For Laboratory, Research, or Manufacturing Use
(1) Snyder, L.R., J. Chromatography, 92,223-230 (1974)
Storage Condition: IMPORTANT: Material will freeze if stored below 17 °C (63°F).

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC



Ken Koehnlein
Sr. Manager, Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700