Methanol
BAKER ANALYZED® LC-MS Reagent
For Use in Liquid Chromatography and Mass
Spectrometry



Material No.: 9830-02

Batch No.: 21I1762009

Manufactured Date: 2021-08-05

Expiration Date: 2023-08-05

Revision No.: 1

Certificate of Analysis

Test	Specification	Result
Assay (CH₃OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Appearance	Passes Test	Passes Test
Gradient Elution Test (a.u.) - 254 nm	≤ 0.01	< 0.01
Fluorescence Trace Impurities, measured as Quinine Base - at 450 nm Emission	≤ 0.3 ppb	< 0.1 ppb
Fluorescence Trace Impurities, measured as Quinine Base - at Emission Maximum for Impurities	≤ 1.0 ppb	0.3 ppb
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Water (H₂O)	≤ 500 ppm	< 500 ppm
LC/MS Suitability - Largest Response on ESI- Positive Mode (as Reserpine)	≤ 50 ppb	< 50 ppb
Trace Impurities - Aluminum (AI)	≤ 50.0 ppb	< 5.0 ppb
Trace Impurities - * Calcium (Ca)	≤ 50.0 ppb	7.0 ppb
Trace Impurities - Iron (Fe)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities - Lithium (Li)	≤ 30.0 ppb	< 1.0 ppb
Trace Impurities - Magnesium (Mg)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities - Nickel (Ni)	≤ 30.0 ppb	< 5.0 ppb
Trace Impurities - Potassium (K)	≤ 50 ppb	< 10 ppb
Trace Impurities - * Sodium (Na)	≤ 50.0 ppb	< 5.0 ppb

Methanol
BAKER ANALYZED® LC-MS Reagent
For Use in Liquid Chromatography and Mass
Spectrometry



Material No.: 9830-02 Batch No.: 2111762009

Test Specification Result

For Laboratory, Research, or Manufacturing Use

* May change over time due to extraction from glass container.

Filtered through a 0.2 micron filter.

Country of Origin: USA

Packaging Site: Phillipsburg Mfg Ctr & DC

