Methanol BAKER ANALYZED[®] LC-MS Reagent For Use in Liquid Chromatography and Mass Spectrometry





Material No.: 9830-02 Batch No.: 21J1562002 Manufactured Date: 2021-09-17 Expiration Date: 2023-09-17 Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH₃OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Appearance	Passes Test	Passes Test
Gradient Elution Test (a.u.) - 254 nm	≤ 0.01	< 0.01
Fluorescence Trace Impurities, measured as Quinine Base - at 450 nm Emission	≤ 0.3 ppb	< 0.1 ppb
Fluorescence Trace Impurities, measured as Quinine Base - at Emission Maximum for Impurities	≤ 1.0 ppb	0.3 ppb
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.4 ppm
Water (H ₂ O)	≤ 500 ppm	< 500 ppm
LC/MS Suitability - Largest Response on ESI- Positive Mode (as Reserpine)	≤ 50 ppb	< 50 ppb
Trace Impurities - Aluminum (Al)	≤ 50.0 ppb	< 5.0 ppb
Trace Impurities - * Calcium (Ca)	≤ 50.0 ppb	2.0 ppb
Trace Impurities - Iron (Fe)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities - Lithium (Li)	≤ 30.0 ppb	< 1.0 ppb
Trace Impurities - Magnesium (Mg)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities - Nickel (Ni)	≤ 30.0 ppb	< 5.0 ppb
Trace Impurities - Potassium (K)	≤ 50 ppb	< 10 ppb
Trace Impurities - * Sodium (Na)	≤ 50.0 ppb	< 5.0 ppb

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700 Avantor Performance Materials, LLC 100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700 Methanol BAKER ANALYZED[®] LC-MS Reagent For Use in Liquid Chromatography and Mass Spectrometry





Material No.: 9830-02 Batch No.: 21J1562002

Test	Specification	Result

For Laboratory,Research,or Manufacturing Use * May change over time due to extraction from glass container. Filtered through a 0.2 micron filter. Country of Origin: USA Packaging Site: Phillipsburg Mfg Ctr & DC

ames Techie Jamie Ethier Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700 Avantor Performance Materials, LLC 100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700

Page 2 of 2