Methanol BAKER ANALYZED® LC-MS Reagent For Use in Liquid Chromatography and Mass Spectrometry





Material No.: 9830-02 Batch No.: 22G2162001 Manufactured Date: 2022-06-13 Expiration Date: 2024-06-12 Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH3OH) (by GC, corrected for water)	\geq 99.9 %	100.0 %
Appearance	Passes Test	Passes Test
Gradient Elution Test (a.u.) - 254 nm	≤ 0.01	< 0.01
Fluorescence Trace Impurities, measuredFluorescence Trace Impurities, measured as Quinine Base – at 450 nm Emission	\leq 0.3 ppb	0.1 ppb
Fluorescence Trace Impurities, measuredFluorescence Trace Impurities, measured as Quinine Base – at Emission Maximum for Impurities	\leq 1.0 ppb	0.4 ppb
Color (APHA)	≤ 10	5
Residue after Evaporation	\leq 1.0 ppm	0.2 ppm
Water (H ₂ O)	≤ 500 ppm	< 500 ppm
LC/MS Suitability – Largest Response on ESI– Positive Mode (as Reserpine)	≤ 50 ppb	< 50 ppb
Trace Impurities – Aluminum (Al)	\leq 50.0 ppb	< 5.0 ppb
Trace Impurities - * Calcium (Ca)	≤ 50.0 ppb	6.0 ppb
Trace Impurities – Iron (Fe)	\leq 50.0 ppb	< 1.0 ppb
Trace Impurities – Lithium (Li)	\leq 30.0 ppb	< 1.0 ppb
Trace Impurities – Magnesium (Mg)	\leq 50.0 ppb	< 1.0 ppb
Trace Impurities – Nickel (Ni)	\leq 30.0 ppb	< 5.0 ppb
Trace Impurities – Potassium (K)	\leq 50 ppb	< 10 ppb
Trace Impurities – * Sodium (Na)	\leq 50.0 ppb	< 5.0 ppb

Methanol BAKER ANALYZED® LC-MS Reagent For Use in Liquid Chromatography and Mass Spectrometry





Material No.: 9830-02 Batch No.: 22G2162001

Test	Specification	Result

For Laboratory,Research,or Manufacturing Use * May change over time due to extraction from glass container. Filtered through a 0.2 micron filter.

Country of Origin: USA Packaging Site: Phillipsburg Mfg Ctr & DC

James Techie

Jamie Ethier Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700 Avantor Performance Materials, LLC 100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700